

OFFICE OF THE DIRECTOR

ACHARYA HARIHAR REGIONAL CANCER CENTRE, CUTTACK

Order No: 337 /AHRCC, Cuttack dated the 21-01-2019
S-40/2015-19

CORRIGENDUM NOTICE

Sub: - Tender notice no. 5672 dt.05.12.2018 for PET-CT on PPP mode and corrigendum notice no. 6096 dt.29.12.2018.

This corrigendum relates to the above tender notice. The pre-bid meeting for PET-CT on PPP mode for AHRCC was held on 21.12.2018. The firms who attended the pre-bid meeting have submitted their requests, suggestions, and objections. After deliberate consideration of the tender committee over various issues raised by the prospective bidders, the amendments made to the original tender paper are as follow:

Pg./Sl. No.	Original Technical Specifications	Queries Raised by Various Prospective Bidders	Amendments of specification in response to the queries
4/6	Bidders shall ensure that their bids complete in all respects, are submitted by Registered Post/Speed Post/Courier only.	Bidders shall ensure that their bids complete in all respects, are submitted by Registered Post/ Speed Post/ Courier/ in person/Drop Box with acknowledgement of the Parcel.	Bidders shall ensure that their bids complete in all respects, are submitted by Registered Post/Speed Post/Courier/By Person with acknowledgement of the parcel.
6/4.1	Past performance of similar services provided in private/public sector in last three years and user's certificate regarding satisfactory completion of such job as per proforma given in "Appendix C"	Past performance record of minimum one year with user's certificate as per proforma provided in Appendix-C.	Past performance of diagnostic imaging services like CT/MRI provided in private/public sector in last three years out of which PET-CT services should not be less than one year and user's certificate regarding satisfactory completion of such job as per proforma given in Appendix- C.

Dr. S. S. S.
20.1.15

6/ 6	Contract Validity Period: The contract with the selected service provider shall be initially signed for a period of "5 years" and the prices quoted shall remain valid for the duration of the contract. The contract may be extended for another term of "5 years" based on performance and mutual consent.	The bidders were of the view that as PET CT is high value equipment and we depend upon financial institution to get financial assistance, we request you to extend the period for 10 years. The contract may be extended for another term of "5 years" based on performance and mutual consent.	No change in contract validity period except price escalation of not more than 3% annually as per pg. 14 clause no 10 (f).
9/11.3a	All PET-CT cases declared as emergency by the referring authority within 12 hours .	This point should be deleted or can be replaced as 'emergency cases should be given on priority.	All PET-CT cases declared as emergency will be done on priority .
9/11.4	The service provider shall also ensure, at its own cost, IT enabled work station at the radiology department of the institute, where the images and soft copy of the report of the patient should reach within the stipulated time.	Bidders requested to delete this point.	Deleted.
9/11.5	Service provider shall deploy the above staffs to run the facility round the clock (24x7x365) .	It should be minimum 8hrs./day. PET-CT Center will remain closed on Holidays and Sundays. It should be 8hrs./day. Holidays and weekends are not in the calculations. We depend on FDG producing companies, carrier etc. It should be 8hrs./day & 5-days a week.	Service provider shall deploy the above staffs to run the facility for minimum 10 working hours on working days .
14/10.d (a)	Penalty Clause: For delay in Installation & Commissioning For delay beyond the above stipulated period of 120 days , a penalty @ 0.1% of the performance security amount for each week of delay shall be deducted from the monthly billing amount.	Bidders requested to amend this period to 180 days, because it depends upon different approval from AERB which is a time taken procedure.	Penalty Clause: For delay in Installation & Commissioning For delay beyond the above stipulated period of 180 days , a penalty @ 0.1% of the performance security amount for each week of delay shall be deducted from the monthly billing amount.
14/10.f	A single price for all types of PET-CT Scan has to be furnished in separate formats as per appendix-G (all types of 16 slice PET-CT scan) .	Bidders requested for rate escalation of 10% year on year on the quoted price.	A single price for all types of PET-CT Scan has to be furnished in separate formats as per appendix-G (all types of 16 slice PET-CT scan). The annual price escalation for PET-CT scan should not be more than 3%.

Done
21.1.15

15/o	The service provider will have to maintain an uptime of 95% with maximum 7 days of downtime at a stretch.	95% uptime should be calculated on daily operating hours which is 8hrs./ day with 12 days of down time at a stretch.	The service provider will have to maintain an uptime of 95% (minimum 10 hours per working day) with maximum 12 days of downtime at a stretch.
16/10.z	The service provider shall be responsible for storage of images and reports of all PET CT scan done by the service provider including image retrieval system as per the extant laws of MCI. Image retention for medico legal cases for a period of 10 years is the responsibility of the service provider.	Image retention should be the responsibility of the hospital authority. The service provider will hand over the soft copy of the image to the authority on quarterly basis.	In medico-legal cases, the images to be retained by both parties for the stipulated period, i.e., 10 years.
18/ Appendi x-B	1. Messrs ----- (name and address of the agent) is our authorized agent for ----- -----	Appendix-B: 1. Need to change as "Authorized Bidder".	Messrs ----- (name and address of the agent) is our authorized bidder for - -----
19/ Appendi x-C	ASSIGNMENT OF PET-CT SCAN SERVICES SUCCESSFULLY COMPLETED DURING LAST THREE YEARS	ASSIGNMENT OF PET-CT SCAN SERVICES SUCCESSFULLY COMPLETED DURING LAST ONE YEAR	ASSIGNMENT OF PET-CT SCAN SERVICES SUCCESSFULLY COMPLETED DURING LAST ONE YEAR
28/B.3	A latest whole body Positron Emission Tomography system should employ non-hygroscopic, high light yield(>/= 80% of NaI) and low decay time scintillator material like LYSO, LSO or equivalent for detecting 511 Kev gamma photons in coincidence. System should have Time-of-flight (TOF) technology	A latest whole-body Positron Emission Tomography system should employ non-hygroscopic, scintillator material like LYSO, LSO, BGO or equivalent for detecting 511 Kev gamma photons in coincidence. System should have Time-of-flight (TOF) technology or other latest reconstruction technology.	A latest whole body Positron Emission Tomography system should employ non-hygroscopic and low decay time scintillator material like LYSO, LSO or equivalent for detecting 511 Kev gamma photons in coincidence. System should have Time-of-flight (TOF) technology or any latest reconstruction technology.
28/B.4	Number of crystals should be 20,000 or more and it should be 4 x 4 x 20 mm or better in size.	Bidders requested to delete this point.	Deleted.
28/B.5	Number of PMTs should be 450 or more.	Bidders requested that number of PMTs should be 425 or more.	Number of PMTs should be 425 or more.
29/10	System should utilize the Time-of-flight (TOF) algorithm during image reconstruction for better image quality at low radiation dose.	Time-of-flight (TOF) technology or other latest reconstruction technology.	System should utilize the Time-of-flight (TOF) algorithm or other latest reconstruction technology for better image quality at low radiation dose.

Done
21.11.15

29/12.d	Scatter fraction of the system should be 34% or less at low activity concentration.	Bidders requested scatter fraction of the system should be 38% or less at low activity concentration.	Scatter fraction of the system should be 38% or less at low activity concentration.
30/E.1.a	System should be provided with ECG machine to acquire gated cardiac PET acquisition.	Bidders requested to delete this point.	Deleted.
30/E.1.b	System should be capable of acquiring respiratory gating with required software and hardware and should be capable to transfer the data to TPS for 4D respiratory getting RT planning.	Bidders requested to delete this point.	Deleted.
30/E.2.a	Processing workstation should be server- client bases with minimum 4 clients and one server.	Processing work station should be server client based with ONE client and one server Two processing workstations will be provided with Oncology based software.	Processing workstation should be server- client based with minimum 2 clients and one server.
31/E.2.f	Quote complete PET cardiac quantification i.e. Emory Cardiac toolbox or Cedar cardiac package or equivalent.	Bidders requested to delete this point.	Deleted.
32/G.1	The complete system should have a guarantee including the radioactive reference source, crystals, detectors and x-ray tubes replacement for a period of two years after the satisfactory commissioning (functional delivery) and handing over of the equipment. Warranty will include all the accessories and third party items.	Bidders requested to delete this point.	Deleted.
32/G.2	Rates for Comprehensive maintenance contract (CMC) for whole system including x- ray tube replacement as and when required and accessories for a period of FIVE years after the expiry of warranty period.	Bidders requested to delete this point.	Deleted.
32/G.3	At least 95% uptime should be maintained during warranty as well as CMC period.	Bidders requested to delete this point.	Deleted.
32/G.4	On-site Training to be imparted to staff	Bidders requested to delete this point.	Deleted.


20/11/18

	identified by the institute for One month		
34/4	Appendix –E (Forwarding Letter for Technical Bid)	Typographical error, i.e. Appendix E instead of F	Appendix –F (Forwarding Letter for Technical Bid)
34/10	Appendix - H (Declaration Affidavit on Stamp Paper)	Typographical error, i.e. Appendix H instead of I	Appendix - I (Declaration Affidavit on Stamp Paper)
34/11	Copies of the annual audited statement/Annual Report for 2012-13, 2013-14, 2014-15 or 2013-14, 2014-15 & 2015-16 (If audited) (Provisional statement of account shall not be considered).	Bidders requested for the audit statement period of 2015-16, 2016-17, 2017-18	Copies of the last 3 years annual audited statement/Annual Report

As requested by the bidders, the undersigned accepted the following:

1. Need to provide structural certificate for the installation site.
2. 150 KVA power should be provided in the point of use.
3. Need to provide DG installation at place nearer to the location.
4. Site should be AERB approved (approval should be in the name of the service provider).
5. Civil modifications as per AERB need to be completed by tender inviting authority.

Hence, the last date of submission of tenders complete in all respects is hereby extended until 13.02.2019 by 5 PM. The date of opening of technical bids is hereby fixed on 14.02.2019 at 12 Noon.


21/1/2019
Director