



OFFICE OF THE DIRECTOR

ACHARYA HARIHAR POST GRADUATE INSTITUTE OF CANCER, CUTTACK

E-mail:- ntstore.ahrcc.2019@gmail.com

No. 4247 /AHPGIC, Cuttack dated 24-09-2026

File No S-05/2026

Sealed quotations are invited from the Manufacturers/Authorized Dealers/Registered Firms for supply of Biosafety Cabinet for Day care unit of this Institute. The quotationer should quote make, model, unit price, guarantee/warranty periods, taxes, delivery period, validity and up-to date GST clearance certificate along with their quotation. The last date of submission of quotation is fixed on date **03.10.2026 at 4 PM**. The undersigned reserve the right to cancel any or all of the quotations without assigning any reason thereof.

SPECIFICATION

- Environment Protection.
- Sound level no more than 65 dbA.
- The Unit is suitable for CLASS 100 conditions.
- Outer body made of Board laminated with sunmica top OR MS Powder Coated OR Stainless Steel 304 Grade.
- 70% of the air is re-circulated through the HEPA filter.
- The contaminated air is drawn by the blower through the plenum of the cabinet.
- 30% of the air is exhausted from work area.
- Filtered Exhaust protects ambient environment, and avoids build up, air borne particles.
- The exhaust of Safety Cabinet is connected to the Virus Burn out Unit for product and personnel protection.
- Air is drawn through Pre-Filter and is made to pass through high effective HEPA (High Efficiency Particular Air) filters.
- HEPA filters efficiency 99.97% for particle size 0.3 microns.
- Inflow velocity remains approx. 105 fpm (0.5 m/sec), while down flow velocity remains approx. 55 fpm (0.3 m/sec). 70% air is re-circulated and 30% exhaust.
- Fitted with Motor-Blower. Complete with HEPA Filters, Pre filters, Fluorescent Tube, UV. Tube, Motor Blower, Transparent Acrylic Sheet Front and Side Doors with 5/15 amp socket.
- Power 220-230 Volts AC Supply.

Working Table Size:

(a) 3' x 2' x 2'

(b) 4' x 2' x 2'

[Signature]
24/09/26

Stores Medical Officer

Memo No. 4248 /dt: 24-09-2026

Copy forwarded to the Nizarat Officer, Collectorate, Cuttack /Principal, SCB Medical College, Cuttack/ Superintendent, SCB Medical College and Hospital, Cuttack for information with a request to hang this in their notice board.

[Signature]

[Signature]
24/09/26

Stores Medical Officer