



OFFICE OF THE DIRECTOR  
ACHARYA HARIHAR POST GRADUATE INSTITUTE OF CANCER  
(Formerly: Acharya Harihar Regional Cancer Centre)  
CUTTACK-753007 (Odisha)

Tel.No: 0671-2304264, Website- [www.ahpgic.in](http://www.ahpgic.in)  
Email: [ntstore.ahrcc.2019@gmail.com](mailto:ntstore.ahrcc.2019@gmail.com), [ahrccdirector@gmail.com](mailto:ahrccdirector@gmail.com)

Tender Notice No. 2121 /AHPGIC, Cuttack  
S-05/2025-26

Date: 07-5-2026

Sealed tenders are invited in two-bid system by the Director, Acharya Harihar Post Graduate Institute of Cancer, Cuttack-753007 (formerly known as Acharya Harihar Regional Cancer Centre) for supply of Medical Consumables, Suture Materials, Anaesthesia Disposables & Surgical Dressings on rate contract basis for the period of two years. The bid document and all details can be downloaded only from the website of AHPGIC, [www.ahpgic.in](http://www.ahpgic.in).


The tender schedule dates are as mentioned below:

Sl. No.	Particulars	Date & Time
1	Last date & time of tender submission (by speed post/registered post) only	09/06/2026 by 5:00 PM
2	Last date & time of receipt of pre-bid queries through mail ( <a href="mailto:ahrccdirector@gmail.com">ahrccdirector@gmail.com</a> / <a href="mailto:ntstore.ahrcc.2019@gmail.com">ntstore.ahrcc.2019@gmail.com</a> )	19/05/2026 by 12.30 PM
3	Date and time of pre-bid meeting	20/05/2026 at 1 PM
4	Date & time of opening of technical bid (in presence of bidders/their authorized agents)	10/06/2026 at 1 PM
5	Date of opening of price bid	To be informed to qualified bidders only

The authority reserves the right to accept/reject any/all bids without assigning any reasons thereof.

In case any modifications/corrections are made to the original tender document after the pre-bid meeting, the same shall be updated in the website of AHPGIC only and no corrigendum in this regard shall be published in newspapers. In case of any bid amendment and/or clarification, it is the sole responsibility of the prospective bidder to download the same from the website of AHPGIC, Cuttack.

In case the last date of submission of bid/date of pre-bid meeting/date of opening of technical bids falls on a holiday, then the next working day shall be the default for the same.

  
Director  
AHPGIC, Cuttack  
07/5/26





OFFICE OF THE DIRECTOR  
ACHARYA HARIHAR POST GRADUATE INSTITUTE OF CANCER  
(Formerly: Acharya Harihar Regional Cancer Centre)  
CUTTACK-753007 (Odisha)  
Tel.No: 0671-2304264, Website- [www.ahpgic.in](http://www.ahpgic.in)  
Email: [ahrccdirector@gmail.com](mailto:ahrccdirector@gmail.com)

Tender Notice No. 2121 /AHPGIC, Cuttack  
S-05/2025-26


Date: 07.05.2026

Sealed tenders in two-bid system are invited from the Manufacturers/Authorized Dealers/Registered firms/EPM or DGSD rate contract holders for supply of medical consumables, Suture materials, Anaesthesia Disposables and surgical dressings on rate contract basis for the period of two years. The tenderers should quote make, model, unit price, taxes, delivery period, and guarantee/ warranty period (if any) and enclose self-attested photocopy of GST registration certificate along with their tender. The last date of submission of tender is fixed on dt 09/06/2026 by 5:00 PM. Last date & time of receipt of pre-bid queries through e-mail ([ahrccdirector@gmail.com](mailto:ahrccdirector@gmail.com)/[ntstore.ahrcc.2019@gmail.com](mailto:ntstore.ahrcc.2019@gmail.com)) is fixed on dt 19/05/2026 by 12.30 PM. **Queries submitted on the Pre-Bid date shall not be entertained under any circumstances.** The date of pre-bid meeting has been fixed on dt 20/05/2026 at 1 PM. The technical bids shall be opened on 10/06/2026 at 1 PM. The tenders must be submitted by speed post/registered post only. The tenderer shall submit EMD of Rs. 2,00,000/- along with tender documents separately. The prices quoted shall be valid for two years. The undersigned reserves the right to cancel any or all bids without assigning any reason thereof. The bidder can bid for one or more lots of items listed as per his/her preference and should mention the lots applied for on top of the envelope.

### **TERMS & CONDITIONS**

1. The tenderers shall submit the tender documents with technical bid in sealed envelope super scribed as technical bid (Cover-A) and price bid in sealed envelope super scribed as price bid (Cover-B), which will be kept inside single sealed envelope super scribed as "Tender for Purchase of Medical Consumable Items" on its top.
2. The tenderers have to submit the following documents along with the tender paper in Technical Bid (Cover-A):
  - a. Self attested Xerox copy of Pan Card in the name of the agency/proprietor.
  - b. Self attested Xerox copy of up to date GST Registration Certificate.
  - c. Self attested copy of up to date GST return certificate.
  - d. Self attested copy of EU/CE/USFDA/ISO compliance certificate should be submitted by bidders quoting for the items for ascertaining quality of the quoted items.
  - e. Self attested Xerox copy of up to date valid Drugs License /Retention copy/Renewal copy.
  - f. Self attested Xerox copy of Non-conviction certification from the Drug Authorities.
  - g. An undertaking that the quoted prices are not more than the prevailing market price/MRP.
  - h. DD of Rs. 5,000.00 (Non-Refundable) against purchase of tender paper in the name of **The Director, Acharya Harihar Regional Cancer Centre, Cuttack.**
  - i. Account payee DD worth Rs. 2,00,000.00 towards cost of EMD in the name of **The Director, Acharya Harihar Regional Cancer Centre, Cuttack.**
  - j. Credentials of last three years towards supply of similar items to any Govt. Hospital/Pvt. Hospital of repute should also be enclosed in the bid document.

- k. Certificate regarding Annual turnover of 3 preceding years, which should be not less than 50 Lakh shall be submitted with the tender documents.
3. The delivery should be at the Non-Technical Store of this hospital in good condition.
4. Any conditional tender will not be accepted.
5. No advance payment will be made against supply of consumables. The required items are to be procured on credit basis against different purchase orders, and payment will be made on submission of bills in triplicate to the undersigned. The authority will not be liable for delay in payment in an unavoidable circumstance like paucity of funds, etc.
6. The rate of the items should be quoted in the Indian currency both in words and figures, excluding taxes. The tax position and other charges if any should be clearly mentioned separately.
7. As this is a tertiary cancer care hospital, these items will be required on emergency basis. The supplier should supply the items as far as possible after the issue of the purchase order taking into the patient care service.
8. The bidder should have their supplying branch office in the Cuttack or Bhubaneswar Municipal Corporation Area to meet the emergency requirement of the hospital. The detailed address and phone number should be mentioned.
9. The rates quoted shall be valid for a period of two years from the date of finalization of the tender.
10. Tenderers qualified in the technical bid shall be called for opening of price bid of the tender, the date and time of which shall be intimated in due course.
11. No tender beyond the stipulated date & time will be accepted, and Tenders without EMD and tender paper cost shall be rejected outright.
12. EMD of the unsuccessful bidders will be returned after finalization of the tender process.
13. Original documents shall be produced for verification if asked for.
14. If any information/documents furnished by the tenderers with the tender paper are found to be misleading or incorrect at any stage, their tenders will be cancelled and the said tenderers shall be barred from taking part in any further tender process in subsequent years & the EMD shall be forfeited.
15. Either the tenderer or his/her authorized representative may remain present at the time of opening of the tenders. However, their absence shall not be a bar for opening of tenders.
16. Tender items are to be quoted in price per unit.
17. If the product is found unsuitable for use after the receipt of the items, the same will be replaced by the supply firm.
18. The selected tenderer shall not discontinue supply during the contract period under any circumstances. Otherwise, the EMD will be forfeited and legal action as deemed proper shall be taken against the selected tenderer.
19. The validity of the tender will remain in force for a period of two years from the date of approval, and on satisfactory performance by the party, it may be extended for another one year at the sole discretion of the authority.
20. The undersigned reserves the full right to accept in full or part or reject any or all the tenders without assigning any reason thereof.
21. Authorized dealership certificate wherever necessary should be attached with the bid documents.



Director  
AHPGIC, Cuttack.



<b>LOT-1</b>		
<b>(USFDA Compliance certificate should be submitted in Technical Bid)</b>		
<b>LIST OF HOSPITAL CONSUMABLES</b>		
<b>( Thermoplastic Masks )</b>		
<b>SL. NO.</b>	<b>DESCRIPTION</b>	<b>SPECIFICATIONS</b>
1	Precut of Head and Neck	4 clamp, 2.4 mm
2	Precut of Head and Neck	3 clamp, 2.4 mm
3	Precut of Hip and Pelvis	4 clamp, 2.4 mm
4	Precut of Pediatric	3 clamp

<b>LOT-2</b>		
<b>(USFDA Compliance certificate should be submitted in Technical Bid)</b>		
<b>LIST OF HOSPITAL CONSUMABLES</b>		
<b>(IV &amp; Oral Contrasts)</b>		
<b>SL. NO.</b>	<b>DESCRIPTION</b>	<b>SPECIFICATIONS</b>
1	Inj. Iopromide-300 mg	50 ml(Vial/Bottle)
2	Inj. Iohexol-300 mg	50 ml(Vial/Bottle)
3	Inj. Diatrizoate Meglumine Diatrizoate Sodium 76%	20 ml(Ampule)



**LOT-3**  
**LIST OF HOSPITAL CONSUMABLES**

<b>SL. NO.</b>	<b>DESCRIPTION</b>	<b>SPECIFICATIONS</b>
1	Black Braided Silk 1	(2 × 90cm)
2	Cytology Needle	20G x 15
3	Surgical needles	Suture needles, curved cutting needle different sizes 6 to 10
4	Tracheostomy tube (Metallic)	Sizes 30, 32 mm
5	Cricothyroidectomy set	Sizes 30, 32 mm
6	Urostomy Bag	
7	Vaccu Canistors	
8	Vaccu Sponge	
9	Pneumatic compression stockinet	Size- Small,Medium,Large
10	Warming blanket (Disposable)	
11	Vascular Tape	
12	J P Drain	6 Fr, 5 Fr
13	Intubating Introducer	14 Fr. Length - 70 cm have a rigid metallic Stylet- length-60 cm
14	Airway Exchange Catheter	Connector for Oxygenator -14 Fr, Length- 83 cm
15	Implantable Port/ Chemo Port	Implantable Access port (Venos). Made up of PSU/Titanium Chamber with Silicon Membrane and suture holes. Catheter Size- 8 Fr. To 10 Fr.. Catheter material- Silicon or Polyurethane. CT Compactable, MRI Conditional, MRI conditional, PVC Free. 2000 Punctures minimum with 22G Non-coring needle & Lase marked radiopaque catheter. <b>(USFDA)</b> Compliance certificate should be submitted. 8 Fr,9.5 Fr,9.6 Fr,10 Fr
16	Implantable Port/ Chemo Port Compatible Needle	20G × 1.0 in. (0.9mm × 25mm) <b>(USFDA)</b> compliance Certificate should be submitted.
<b>The L1 Bidder for Chemoport Needle will be considered as L1 bidder for Both Needle &amp; Chemoport.</b>		
17	Pediatric Ventilator Circuit	
18	Ryle's Tube	No.- 9.0
19	Feeding Tube	No. 10
20	Adhesive elastic bandage	6 inches
21	Vascular clip (Size -100)	Vascular clip (Cartridge) Reusable clip applicator (for open surgery) <b>(USFDA)</b> compliance Certificate should be submitted
22	Vascular clip (Size -200)	Vascular clip(Cartridge) Reusable clip applicator (for open surgery) <b>(USFDA)</b> compliance Certificate should be submitted
<b>The L1 bidder for catridge will be considered as L1 bidder for both clip and clip Applicator.</b>		
23	Cover slip (Glass)	Square 22x22mm, 18x18 mm; Rectangular 22x30mm, 22x40mm
24	Needle syringe destroyer	metal plastic

25	LMA	Supraglottic airway device of Silicone material, 15 mm connector. Should have drain tube, bite block and wire reinforced airway tube.
26	Rebreathing bag	500 ml, 2 ltr
27	IV Cannula	Adult – two way with closing Cover Type II sterilized, sizes 14, 18, 20, 22 having radio opaque catheter with CE certification. Child – two way with closing Cover Type II sterilized, size 24 (child) having radio opaque catheter with CE certification.
28	T piece	Flexible corrugated tube made up of medical grade PE. Mount connects to ET tube. Free moving swivel. Self sealing suction port. Size: Universal
29	Sterilized surgical suture black braided silk	Size 1-0, length 2x75 cm (Precut and presterilized) U.S.P. with CE certification
30	Ryle's tube with integrated guide wire (freka tube)	No. 14, 16 & 18
31	Whitacare Spinal Needle	Size- 25 ,26 ,27
32	Tube Exchanger	
33	Tegaderm	
34	Sodalime	5 kg jar
35	HME Filter	Made up high strength plastic material, high efficiency electrostatic filters. Sizes: Adult Sizes: Paediatric
36	C.V.P Measuring Device	Manometer tube with graduation from -4 cm to +34 cm and attached to three way tap, provided with 120 cm long extension tube , fitted with female luer lock on one end and male luer lock on other end , Latex free injection site sliding indicator , moulded clamps with latex free rubber string
37	Urinary Drainage Bag	Sterilized with non-return valve and Drainage outlet with a capacity of 2000 ml with marking, nontoxic. Pyrogen free, double seek, clinical grade PVC.
38	Scalp Vein	Sizes- 20
39	Scalp Vein	Sizes- 22
40	Scalp Vein	Sizes- 26
41	Ryle's Tube	Medical grade PVC low friction, universal funnel shaped connector, distal end-conned with corrosion resistant stainless steel balls. Sizes -10, 12,14,16,18,20
42	Corrugated Drainage Sheet	
43	Infant feeding tube	Sizes 3, 4, 5, 6, 7, 8

44	Central Line	Manufactured from specially formulated and biocompatible PU material, Polyurethane anti kink soft tip, radio-opaque catheter, soft & flexible J-tip guide wire with Y introducer needle, Kit Should contain needle free connector														
		<table border="1"> <thead> <tr> <th>Type</th> <th>Size</th> </tr> </thead> <tbody> <tr> <td>Double Lumen</td> <td>Fr 5 × 8 cm</td> </tr> <tr> <td>Triple Lumen</td> <td>Fr 7 × 16, 20 cm</td> </tr> </tbody> </table>	Type	Size	Double Lumen	Fr 5 × 8 cm	Triple Lumen	Fr 7 × 16, 20 cm								
		Type	Size													
		Double Lumen	Fr 5 × 8 cm													
Triple Lumen	Fr 7 × 16, 20 cm															
45	Micropore for dressing	2 inches														
46	Micro infusion set															
47	Laryngeal Mask	100 % Silicone mask, with special port to provide tracheal ventilation along with gastric access.														
		<table border="1"> <thead> <tr> <th>Size</th> <th>Gastric tube</th> </tr> </thead> <tbody> <tr> <td>1.0</td> <td>4, 5 FG</td> </tr> <tr> <td>2.0</td> <td>6, 7 FG</td> </tr> <tr> <td>2.5</td> <td>8, 9 FG</td> </tr> <tr> <td>3.0</td> <td>10, 12 FG</td> </tr> <tr> <td>4.0</td> <td>12, 14 FG</td> </tr> <tr> <td>5.0</td> <td>14, 16 FG</td> </tr> </tbody> </table>	Size	Gastric tube	1.0	4, 5 FG	2.0	6, 7 FG	2.5	8, 9 FG	3.0	10, 12 FG	4.0	12, 14 FG	5.0	14, 16 FG
		Size	Gastric tube													
		1.0	4, 5 FG													
		2.0	6, 7 FG													
		2.5	8, 9 FG													
		3.0	10, 12 FG													
4.0	12, 14 FG															
5.0	14, 16 FG															
48	Nasopharyngeal Airway	Medical grade PVC soft , flexible and kink resistant, atraumatic, soft rounded edges and smooth surface. Removable 15 mm connector.														
		<table border="1"> <thead> <tr> <th>Size</th> <th>Length</th> </tr> </thead> <tbody> <tr> <td>6.5 FG</td> <td>105 mm</td> </tr> <tr> <td>7 FG</td> <td>125 mm</td> </tr> <tr> <td>7.5 FG</td> <td>130 mm</td> </tr> <tr> <td>8 FG</td> <td>135 mm</td> </tr> <tr> <td>8.5 FG</td> <td>145 mm</td> </tr> </tbody> </table>	Size	Length	6.5 FG	105 mm	7 FG	125 mm	7.5 FG	130 mm	8 FG	135 mm	8.5 FG	145 mm		
		Size	Length													
		6.5 FG	105 mm													
		7 FG	125 mm													
		7.5 FG	130 mm													
8 FG	135 mm															
8.5 FG	145 mm															
49	Tracheostomy tube (cuffed tube)	PVC. Sizes 6.5, 7, 7.5, 8, 8.5, 9														
50	Catgut Chromic Atraumatic	Size 1-0, length 76 cm, ½ circle round bodied (45-mm needle) U.S.P. with CE certification (not mandatory)														
51	Polyglactin 1/0	45 cm 23 mm 1/2 circle reverse cutting														
52	Polyglactin 910	23 mm 1/2 circle reverse cutting (port closure)														
53	Polyglactin 4-0	90cm 25mm needle, ½ circle RB														
54	Polydioxanone	3-0, 45 cm, 13 mm needle ½ circle round bodied														
55	IV infusion Set	Nontoxic And Transparent PVC Moulded Drip Chamber, Disc Type Fluid Filter, Super Smooth Kink Resistant Tubing.														
		Tube Length 150 cm. Filter size 20 drops/ml														
56	Percutaneous tracheostomy kit	Romo soft percutaneous tracheostomy tube , Dilating forcep , J – tip guide wire , Syringe - 10 ml Dilating cannula, Dilator ,Scalpel , Cotton tape.														
		<table border="1"> <thead> <tr> <th>I.D.</th> <th>O.D.</th> </tr> </thead> <tbody> <tr> <td>7.0 mm</td> <td>10.50 mm</td> </tr> <tr> <td>7.5 mm</td> <td>11.30 mm</td> </tr> <tr> <td>8.0 mm</td> <td>11.90 mm</td> </tr> <tr> <td>8.5 mm</td> <td>12.60 mm</td> </tr> </tbody> </table>	I.D.	O.D.	7.0 mm	10.50 mm	7.5 mm	11.30 mm	8.0 mm	11.90 mm	8.5 mm	12.60 mm				
		I.D.	O.D.													
		7.0 mm	10.50 mm													
		7.5 mm	11.30 mm													
8.0 mm	11.90 mm															
8.5 mm	12.60 mm															

57	Bone wax	
58	Mask With Air Seal	
59	J P Drain	
60	Enteral Feeding Tube	1.2 Ltrs., DEHP free Medical Grade PVC with sterile pack
61	Tracheostomy tube with Inner Cannula	Outer Tube (Cuff + Pilot Ballon) Inner tube-2-3 one fenestrated One nonfenestrated size- 7,7.5,8
62	Pipelle Endometrial Biopsy Suction Curette	
63	Malecot Nephrostomy Drainage Catheter	Size:- 10Fr, 12Fr, 14Fr, 16Fr, 18Fr, 20Fr, 22Fr, 24Fr, 26Fr, 28Fr (Length:- 30Cms)
64	Basiloid Extra	5 Ltr
65	Bacillo- 25 spray (Each 100 gm contain to propanol 9 gm, 1 propanol 6gm ethanol 10gm.)	
66	Terumo Guide Wire (Straight tip) 0.32/0.35	
67	Sonoplex Needle	Length- 35mm,40mm,50mm,70mm,80mm,100mm,120mm Gauge-20G,21G,22G,23G,24G,25G Bevel Types-20° or 30°
68	Bone Marrow Biopsy Needle	Cannula should be made of A1S1 304 tempered stainless steel
		· Disposal Device.
		· Should have an angled tip.
		· Should have light ergonomic handle.
		· Should have hammerable top with wide surface area allowing for use of more force.
		· Should have lock allowing aspiration to be performed at any step of the procedures.
		· Should have length 100mm.
69	Bone Marrow Biopsy Needle with external catching system	· Gauge – 13G
		· Disposable device.
		· Should have Ergonomic T-handle with sharpened luer lock allowing aspiration to be performed at any time.
		· Should have fish mouth sharpening on needle tip for good penetration capacity and perfect cylindrical cutting.
		· Should be equipped with extraction wire to measure and remove specimen.
		· Should have external claw system to enable tissue extraction without luxation.
		· Should have length of 100mm.
· Gauge – 11G		



LOT-4

(USFDA Compliance certificate should be submitted in Technical Bid)

LIST OF HOSPITAL CONSUMABLES  
(STAPLERS)

SL. NO.	DESCRIPTION	SPECIFICATIONS	REMARKS
1	Gastro Intestinal Anastomosis Stapler	55mm/60mm for <b>Medium Thick</b> tissue ranges, 3 staggered rows of staplers with open staple height of 2mm-4mm and closed staple height 0.75-2.5mm along with the fresh knife for open surgery	The L1 bidder for cartridge will be considered as L1 bidder for both cartridge and stapler.
2	Cartridge/reloads for gastro Intestinal Anastomosis Stapler		
3	Gastro Intestinal Anastomosis Stapler	55mm/60mm for <b>Extra Thick</b> tissue ranges, 3 staggered rows of staplers with open staple height of 2mm-4mm and closed staple height 0.75-2.5mm along with the fresh knife for open surgery	
4	Cartridge/reloads for gastro Intestinal Anastomosis Stapler		
5	Gastro Intestinal Anastomosis Stapler	75mm/80mm for <b>Medium Thick</b> tissue ranges, 3 staggered rows of staplers with open staple height of 2mm-4mm and closed staple height 0.75-2.5mm along with the fresh knife for open surgery	
6	Cartridge/reloads for gastro Intestinal Anastomosis Stapler		
7	Gastro Intestinal Anastomosis Stapler	75mm/80mm for <b>Extra Thick</b> tissue ranges, 3 staggered rows of staplers with open staple height of 2mm-4mm and closed staple height 0.75-2.5mm along with the fresh knife for open surgery	
8	Cartridge/reloads for gastro Intestinal Anastomosis Stapler		
9	Three rows end to end anastomotic stapler	<b>Medium thick</b> tissues with 3 staggered rows of staplers <b>Sizes- 25mm,28mm,29mm,31mm &amp; 33 mm</b>	
10	Three rows end to end anastomotic stapler	<b>Extra thick</b> tissues with 3 staggered rows of staplers <b>Sizes- 25mm,28mm,29mm,31mm &amp; 33 mm</b>	
11	Linear Stapler	<b>30mm</b> for low/ultra low anterior resection	
12	Cartridge for above Linear Stapler		
13	Linear Stapler	<b>60mm</b> for low/ultra low anterior resection	
14	Cartridge for above Linear Stapler		
15	Linear Stapler	<b>45mm</b> for low/ultra low anterior resection	
16	Cartridge for above Linear Stapler		
17	Linear Stapler	<b>90mm</b> for low/ultra low anterior resection	
18	Cartridge for above Linear Stapler		
19	Circular stapler	<b>Sizes- 21, 25,28,29, 31 &amp; 33mm</b>	
20	Disposable Skin Stapler	35mm wide Titanium Stapler	
21	Disposable Skin Stapler remover		

LOT-5 (USFDA Compliance certificate should be submitted in Technical Bid)

LIST OF HOSPITAL CONSUMABLES  
(BIOPSY GUNS)

SL. NO.	DESCRIPTION	SPECIFICATIONS
1	Disposable Automatic Spring Loaded Biopsy Guns	<ul style="list-style-type: none"> <li>• USFDA Approval</li> <li>* Instrument should have choice of two firing buttons. One on the side and other on the rear end. Should have penetration depth of 22mm.</li> </ul> <p><u>Size</u></p> <p>14G 10CM</p> <p>14G 16CM</p> <p>16G 10CM</p> <p>16G 16CM</p> <p>18G 10CM</p> <p>18G 16CM</p> <p>18G 20CM</p> <p>18G 25CM</p> <p>20G 10CM</p> <p>20G 16CM</p> <p>20G 20CM</p>
2	Disposable Semiautomatic Co Axial Biopsy Guns	<ul style="list-style-type: none"> <li>• Two Penetration Depth in a Single Gun (10mm, 20mm)</li> <li>· Convenient adjustable throw of 10mm and 20mm with blunt tip stylet.</li> <li>· USFDA Approved.</li> </ul> <p><u>Size</u></p> <p>18G &amp; 10cm needle length, Coaxial length is 17G x7.8cm</p> <p>18G &amp; 16cm needle length, Coaxial length is 17G x13.8cm</p> <p>20G &amp; 10cm needle length, Coaxial length is 20G x7.8cm</p> <p>20G &amp; 16cm needle length, Coaxial length is 17G x13.8cm</p> <p>16G – 10cm &amp; 16cm</p> <p>18G – 10cm &amp; 16cm</p>
3	Disposable Breast Biopsy Guns	<ul style="list-style-type: none"> <li>• Two throw length – 18mm &amp; 25mm in a single Gun</li> <li>· Trocar tip for easy penetration, Automatic and Semiautomatic throw options</li> </ul> <p><u>Size</u></p> <p>14G – 10cm</p> <p>16G – 10cm</p>